

CE **Sigrist-Photometer AG**
Hofurlistrasse 1
CH-6373 Ennetbürgen

declares under its sole responsibility that the type of instrument **AquaScat S** complies with the essential requirements of the directives and standards listed below.

Directive

2014/53/EU:	Radio Equipment Directive
2014/30/EU:	Electromagnetic compatibility
2011/65/EU:	RoHS

Standards

EN 61326-1:2013	EMC requirements for measuring devices
<i>ETSI EN 301 489-1 V1.9.2</i>	<i>EMC for radio equipment</i>
<i>ETSI EN 301 489-17 V2.2.1</i>	<i>EMC for radio equipment</i>
<i>ETSI EN 300 328 V1.8.1</i>	<i>Broadband transfer systems</i>

Only with integrated WLAN module

Comments: The design, manufacture and testing of the pressure equipment is carried out in accordance with the Pressure Equipment Directive 2014/68/EU Article 4, Paragraph 3 in line with good engineering practice based on the AD 2000 regulations.

Signed for and on behalf of:

CH-6373 Ennetbürgen, 2022-10-25



Marc Achermann, Director R&D



Matthias Schulthess, Director
Product Management