

CE **Sigrist-Photometer AG**
Hofurlistrasse 1
CH-6373 Ennetbürgen

declares under its sole responsibility that the type of instrument **ColorPlus Ex** complies with the essential requirements of the directives and standards listed below.

Directive

2014/30/EU: Electromagnetic compatibility
2014/34/EU: Equipment and protective systems intended for use in explosion hazard areas
Electromagnetic compatibility

Standards

EN 61326-1:2013
EN IEC 60079-0:2018
EN 60079-1:2014
EN 60079-26:2015
EN 61010-1:2010

The following appointed body has carried out the conformity assessment procedure according to Directive 2014/34/EU Annex II:

DEKRA Testing and Certification GmbH
0158
Dinnendahlstrasse 9
DE 44809 Bochum

BVS 10 ATEX E 102 X
IECEX BVS 13.0058X


The following appointed body has carried out the assessment of the module "Production quality assurance" according to Directive 2014/34/EU Annex IV and VII:

DEKRA Testing and Certification GmbH
0158
Dinnendahlstrasse 9
DE 44809 Bochum


BVS 20 ATEX ZQS/E422

Comments: The design, manufacture and testing of the pressure equipment is carried out in accordance with the Pressure Equipment Directive 2014/68/EU Article 4, Paragraph 3 in line with good engineering practice based on the AD 2000 regulations.

Signed for and on behalf of:
CH-6373 Ennetbürgen, 2022-11-10



Marc Achermann, Director R&D



Matthias Schulthess, Director
Product Management