



Ref. Certif. No.

SE-93565

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

| | |
|---|--|
| Product | Tunable Diode Laser Spectrometer |
| Name and address of the applicant | NEO Monitors AS Prost Stabells vei 22, N-2019 Skedsmokorset, Norway |
| Name and address of the manufacturer | Same as applicant |
| Name and address of the factory <i>Note: When more than one factory, please report on page 2</i> | Same as applicant |
| Ratings and principal characteristics | LaserGas™ II SP Compact O2: 18 – 36 VDC, Max. 20 W LaserGas™ Q ICL: 24 VDC, Max. 20 W |
| Trademark (if any) | LaserGas™ |
| Customer's Testing Facility (CTF) Stage used | - |
| Model / Type Ref. | LaserGas™ II SP Compact O2, LaserGas™ Q ICL |
| Additional information (if necessary may also be reported on page 2) | Only hazards resulting from laser radiation have been addressed. |
| A sample of the product was tested and found to be in conformity with | IEC 60825-1:2014 |
| As shown in the Test Report Ref. No. which forms part of this Certificate | MTt9P00102-2 |

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 18 March 2019

intertek

Signature:

Leif Mattsson