



Ref. Certif. No.
SE-92903

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Tunable Diode Laser Spectrometer
Name and address of the applicant	NEO Monitors AS Prost Stabels vei 22, N-2019 Skedsmokorset, Norway
Name and address of the manufacturer	Same as applicant
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	Same as applicant
Ratings and principal characteristics	iQ ² : 18 – 30 VDC, Max. 30 W II SP H2: 18 – 36 VDC, Max. 20 W III SP O2: 18 – 36 VDC, Max. 20 W Q ICL: 18 – 36 VDC, Max. 20 W
Trademark (if any)	LaserGas
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	iQ ² , II SP H2, III SP O2, Q ICL
Additional information (if necessary may also be reported on page 2)	Only hazards resulting from laser radiation have been addressed
A sample of the product was tested and found to be in conformity with	IEC 60825-1:2014
As shown in the Test Report Ref. No. which forms part of this Certificate	MTt8P05601-2

This CB Test Certificate is issued by the National Certification Body

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00

Date: 20 December 2018



Signature:
Paul Klemets